

# Software for Dose Registry – practical aspects

Seminar of the Federal Commission on Radiation  
Protection (KSR) – Radiation Protection in Medicine

**Thiago V. M. Lima PhD**

Medical Physics Expert

Institute for Radiology and Nuclear Medicine

Luzerner Kantonsspital

29.01.2021


# Content

- Introduction and Legislation
- Dose registry software (Dose Management Systems) requirements
  - Common Features
  - Vendor Specific Features
- National Registries
  - Success stories
    - American College of Radiology (ACR) - NRDR
    - Swiss National CT Dose Registry Project
- Conclusion

# Introduction

Typical questions about software for dose registry are:

- Am I legally obligated to have one?
- What does it actually do?







## What is not included in my talk:

- The answer if you should or not buy a software (there are free options)
- Or even which software to get

# Swiss Legislation

## Radiation Protection Ordinance (StSV/ORaP)



814.501

### Strahlenschutzverordnung (StSV)

vom 26. April 2017 (Stand am 1. Januar 2021)

---

*Der Schweizerische Bundesrat,*  
gestützt auf das Strahlenschutzgesetz vom 22. März 1991<sup>1</sup> (StSG)  
und auf Artikel 83 des Bundesgesetzes vom 20. März 1981<sup>2</sup> über die  
Unfallversicherung,  
*verordnet:*

**1. Titel: Allgemeine Bestimmungen**  
**1. Kapitel: Gegenstand, Geltungsbereich und Begriffe**

**Art. 1**                   Gegenstand und Geltungsbereich

<sup>1</sup> Diese Verordnung regelt zum Schutz des Menschen und der Umwelt vor ionisierender Strahlung:

- a. für geplante Expositionssituationen:
  1. die Bewilligungen,
  2. die Exposition der Bevölkerung,
  3. nicht gerechtfertigte Tätigkeiten,
  4. die medizinische Exposition,
  5. die berufliche Exposition.

# Bases of Radiation Protection




# Bases of Radiation Protection




## RPO Art 5 Dose limits:

For planned exposure situations, limits shall be specified which must not be exceeded by the sum of all radiation doses accumulated by a person in a calendar year (dose limit). For medical exposures, no such limits shall be specified.

# Bases of Radiation Protection



# Bases of Radiation Protection



## Section 3 Medical Optimisation Art. 32 Optimisation of medical exposures

<sup>1</sup> In diagnostic, interventional radiology and nuclear medicine examinations, **the licence holder must keep all radiation doses as low as achievable consistent with obtaining the required imaging information.**

<sup>2</sup> For all therapeutic exposures, the licence holder must carry out individual dosimetric planning. The doses for risk organs must be kept as low as is achievable and consistent with the intended radiotherapeutic purpose.

# Bases of Radiation Protection




<sup>3</sup> For the **protection of patients, the optimisation process involves in particular:**

- a. selection of suitable equipment, including software;
- b. consistent production of adequate diagnostic information or therapeutic outcomes;
- c. the practical aspects of procedures;
- d. quality assurance;
- e. assessment and evaluation of the patient dose or the administered activity;**
- f. use of appropriate set-up parameters or appropriate radionuclides;
- g. use of sensitive detectors;
- h. for every medical installation, use of the elements required for the protection of patients.

<sup>4</sup> **The dose received by personnel must be taken into account in the optimisation process.**

<sup>5</sup> The FDHA may issue technical optimisation provisions for the protection of patients.

# Bases of Radiation Protection



## Art. 33 Documentation obligation

**The licence holder must document** all therapeutic and diagnostic exposures in the medium-dose or high-dose range and in mammography in such a way that the **radiation dose received by the patient can be determined at a later date.**


## Art. 35 Diagnostic reference levels

<sup>1</sup> The FOPH shall publish recommendations on radiation doses for diagnostic, interventional or nuclear medicine examinations in the form of diagnostic reference levels.

<sup>2</sup> To this end, it shall conduct national surveys based on the data specified in Article 34 paragraph 2, take international recommendations into account and publish the results.

<sup>3</sup> **License holders must regularly review their own practices and account for any deviations from diagnostic reference levels.**

# Bases of Radiation Protection



In summary, having such registry software IS NOT legally required.

## Art. 33 Documentation obligation

**The licence holder must document** all therapeutic and diagnostic exposures in the medium-dose or high-dose range and in mammography in such a way that the **radiation dose received by the patient can be determined at a later date.**

## Levels

Recommendations on radiation  
al or nuclear medicine  
diagnostic reference levels.



national surveys based on the

data specified in Article 34 paragraph 2, take international recommendations into account and publish the results.

**<sup>3</sup> Licence holders must regularly review their own practices and account for any deviations from diagnostic reference levels.**

# Dose Evaluation and Optimisation

Histogram for 10 PET Tumor Körperstamm



Interne Protokoll  
BAG ref. Protokoll  
Untersuchungen von  
bis  
Anzahl Studien

	CT TK
	10 PET Tumor Körperstamm
	20201001
	20201120
	127


	Mean CTDIvol [mGy]	DLP [mGycm]
Minimum	0.62	45.1
25. Perz	1.32	163.6
Median	1.69	208.5
75. Perz	2.02	253.2
Maximum	3.68	518.9

Zielgrösse BAG	3	200
DRW BAG	6	620

Can you do it without the software? Yes!

# You can even do very fancy stuff

Evolution of Abdomen Standard




Lima, Thiago; Schindera, Sebastian; Scheidegger, Stephan; Lutters, Gerd, 2018. Connecting the missing piece: a retrospective evaluation of image quality and dose in respect to the parameters variability for a clinical CT protocol ECR 2018, EuroSafe Imaging, Vienna, 28 February - 4 March 2018.

...but...,

Individual dose information is stored in a PACS, systematic and comprehensive analysis usually requires high manual effort. But such an analysis of ionising radiation is inevitable for **improvements in radiological processes**.

- Medical Physicist, Radiologist/Nuclear Medicine Physician and Radiographer (MTRA/TRM) resources are limited
- And by using software registries, the local teams are able to streamline/ optimise their workflow



Disclaimer: There are different dose registry software, some are commercial tools others are freely available. During this presentation I will mention the name and features of different solutions, but neither myself nor my employer is financially supported by these solutions.

# Radiation Dose Management System – requirements and recommendation

- Dose management systems (DMS) are **recommended but not mandatory** for X-ray equipment users, to comply with new radiation protection requirements like the EU-BSS.
- DMS are designed for medical practitioners, radiographers, medical physics experts (MPE) and other health professionals involved in imaging to **support their tasks and duties of radiation protection** in accordance with local and national requirements.
- In particular, the requirement according to the ALARA principle is to perform X-ray examinations with the aim to achieve a minimum dose level but maintaining a sufficient image quality or diagnostic accuracy, for the clinical indication.

European Radiology  
https://doi.org/10.1007/s00330-020-07290-x

RADIOLOGICAL EDUCATION



## Radiation dose management systems—requirements and recommendations for users from the ESR EuroSafe Imaging initiative

Reinhard W. Loose<sup>1,2</sup> · Eliseo Vano<sup>3</sup> · Peter Mildenerger<sup>4</sup> · Virginia Tsapaki<sup>5</sup> · Davide Caramella<sup>6</sup> · Johan Sjöberg<sup>7</sup> · Graciano Paulo<sup>8</sup> · Alberto Torresin<sup>9</sup> · Sebastian Schindera<sup>10</sup> · Guy Frija<sup>11</sup> · John Damilakis<sup>12</sup> · on behalf of the European Society of Radiology (ESR)<sup>13</sup>

Received: 21 August 2020 / Revised: 7 September 2020 / Accepted: 11 September 2020  
© The Author(s) 2020

### Abstract

The European Directive 2013/59/Euratom on basic safety standards for protecting people against ionising radiation, and the associated quality assurance. The EuroSafe Imaging initiative, a joint effort of the European Society of Radiology (ESR) and the European Society of Medical Radiology (ESMR), has launched a campaign of the European Society of Radiology (ESR) “EuroSafe Imaging Initiative” with the aim to provide European recommendations on the use of Dose Management Systems (DMS) in clinical practice. The WG follows the “Action 4: “Promote dose management systems” of the EuroSafe Imaging Call for Action. The WG consists of medical physicists, radiographers, medical physics experts (MPE) and other health professionals. The WG has defined the duties of radiation protection in accordance with local and national requirements. The WG has identified the key points when installing a DMS and classified the individual functions at different levels of complexity.


### Key Points

- DMS are very helpful software tools for ensuring patient safety, optimisation, compliance with DRLs and quality assurance.
- DMS can help to fulfil dosimetric aspects of the European Directive 2013/59/Euratom.
- The EuroSafe WG analyses DMS requirements and gives recommendations for users.


**Keywords** Dose management systems · Radiation protection · Optimisation · Quality assurance

+ my personal findings


# Clinical Radiation Protection tasks supported by the DMS



# Typical set-up of the DMS within IT infrastructure




# Data Requirements




# Basic Requirements

These systems should use physical device-related DICOM dose parameters



Store dose data in a database

Local procedure names should be translated into standardised procedure names



Set alert trigger levels (local and national).

Export dose data for optimisation, QA, reporting (e.g. to national authority) and post processing



# Typical features


- Analysis of different dosimetric parameters (CTDIvol, DLP, DAP, ...) and/versus device.



Philips - DoseWise

# Typical features


- Analysis of different dosimetric parameters (CTDIvol, DLP, DAP, ...) and/versus protocol.



OpenREM

# Typical features


- Peak skin-dose map



Qaelum - Dose

# Typical features


- Trend analysis



Qaelum - Dose

# Typical features

- Contrast media



Qaelum - Dose

# Typical features

- CT Organ dosimetry (including uterus doses RPO, art 40)

The screenshot displays the Radimetrics software interface for CT Organ dosimetry. The top navigation bar includes 'EXAM DETAILS', 'INTERACTIVE DOSIMETRY', 'LOGBOOK', 'RDSR', and 'DOSE SHEET'. The main interface is divided into several sections:

- Scan Demographics:** Age: 55 y, Gender: Male, Weight: Diameter: 251 mm, Phantom: Adult Male <294 mm. Buttons for 'Select Phantom' and 'Reset' are present.
- Acquisitions:** (Empty section)
- Simulation Parameters:**
  - Scan Region: (Dropdown menu)
  - Slice Thickness: (Input field)
  - Tube Voltage: (Input field)
  - Phantom: (Input field)
  - Overranging Length: (Input field)
  - Exposure per Slice: (Input field)
  - Collimation: (Input field)
  - Pitch: (Input field)
  - Scan Range: (Input field)
  - Scan Length: (Input field)
  - CTDIvol: (Input field)
  - nCTDIvol: (Input field)
  - 'Reset' button
- Organ Dose:** A horizontal bar chart showing organ doses. A 'Run Simulation' button is at the top right. The chart lists various organs and their corresponding dose values.

Organ	Dose
Brain (CT)	0.0
Salivary Glands (CT)	0.0
Esophagus (CT)	6.3
Thyroid (CT)	0.1
Lungs (CT)	5.0
Thymus (CT)	0.7
Heart (CT)	6.4
Liver (CT)	22.4
Spleen (CT)	22.1
Stomach (CT)	23.9
Gall Bladder (CT)	22.5
Pancreas (CT)	19.4
Adrenals (CT)	20.2
Kidneys (CT)	26.1
Colon (CT)	22.0
Small Intestine (CT)	22.1
Urinary Bladder (CT)	25.3
Testicles (CT)	17.7
Skin (CT)	8.4
Muscle (CT)	11.4
Red Marrow (CT)	10.5
Skeleton (CT)	14.3
Remainder ICRP103	11.8
Remainder ICRP60	11.8

Bayer - Radimetrics

# Vendor specific features

- GE Dosewatch

Patient dosimetry based on phantom alignment by scout/topo analysis.

## M85101DEED Dosewatch Duke Organ Dose (Organdosis)

### Überblick:

Aktiviert das Organ-Dosis-Modul für CT unter Verwendung der lizenzierten Dosimetrietechnologie der Duke University.

Unter Verwendung der XCAT-Phantombibliothek der Duke University, auf die in über 100 von Experten geprüften Artikeln verwiesen wird, verwendet das Modul einen einzigartigen auf Faltungen basierenden Ansatz für die Organdosis, einschliesslich der Anzeige von Konfidenzintervallen.

Dieses Modul verwendet mehr als 160 XCAT-Phantome, berücksichtigt die Röhrenstrommodulation und bietet eine Schätzung der Dosis für Gewebe ausserhalb des Sichtfelds.

Das Modul bietet auch eine Schätzung der fetalen Dosis basierend auf der Schwangerschaftswoche und der Verwendung von 50 XCAT-Phantomen für schwangere Patientinnen.


Verfügbar für DoseWatch 3.x-Systeme.



# Vendor specific features

- Siemens Teamplay

No local server installation and protocol mapping tool



# Vendor specific features

- Qaelum Dose

Quantitative image quality scoring (Global Noise Level)




Figure 1: Schematic overview of the noise detection algorithm from Christianson et al. (2015). A) Original Image, B) Noise map of soft tissue, C) Histogram of soft tissue noise map.



## Vendor specific features

- Philips DoseWise (plus DoseAware)


Integration with staff doses



**PHILIPS**

This was just at the local level

# National Dose Registry




The **purpose** of the National Dose Registry is to **serve as a central platform to collect, analyse and publish dose information** from different radiology institutes. Its main goals are:

- **Reduce the effort** of participating institutes by offering a semi-automatic process.
- Build a larger and more up-to-date dataset to **improve the quality** of the evaluation results.
- Allow institute benchmarks, leading to **harmonisation of radiation doses**.
- More **frequent publication** of diagnostic reference levels.


# ACR – National Radiology Data Registry

- The primary purpose of NRDR is to **aid facilities with their quality improvement programs and efforts to improve patient care by comparing facility data to that of their region and the nation.** A practice or facility may choose to participate in any or all registries as appropriate for their practice.




# A success history – DIR (Dose Index Registry) from ACR

DIR Facilities  
Jan-Jun 2020



Total 2433 facilities submitted data for 2020Q1Q2 (indicated by black dots)



2433 Facilities connected!!!

# Outcomes – DRL based on 10 most used procedures

Exam Name	Median Patient Size	CTDIvol (mGy)		SSDE (mGy)		DLP (mGy-cm)	
		DRL	AD	DRL	AD	DRL	AD
Head and brain without contrast	14-16	56	49	.	.	962	811
Neck with contrast	18-22	19	15	.	.	563	429
Cervical spine without contrast	18-22	28	20	.	.	562	421
Chest without contrast	29-33	12	9	15	11	443	334
Chest with contrast	29-33	13	10	15	11	469	353
Chest pulmonary arteries with contrast	29-33	14	11	17	13	445	357
Abdomen and pelvis without contrast	29-33	16	13	19	15	781	639
Abdomen and pelvis with contrast	29-33	15	12	18	15	755	608
Abdomen, pelvis and kidney without contrast	29-33	15	12	19	14	705	576
Chest, abdomen and pelvis with contrast	29-33	15	12	18	14	947	779

## Radiation Units in Computed Tomography

Term	Description	Unit
CT Dose Index (CTDIvol)	Radiation energy absorbed per unit mass; for CT, determined for a standard phantom and not a patient	gray (Gy) or milligray (mGy)
Dose Length Product (DLP)	Absorbed dose multiplied by the length of exposure; for CT, determined for a standard phantom and not a patient	milligray-cm (mGy-cm)
Size Specific Dose Estimate	A patient dose estimate which takes into consideration corrections based on the size of the patient	milligray (mGy)

1. Kanal KM, Butler PF, Sengupta D, et al. U.S. Diagnostic Reference Levels and Achievable Doses for 10 Adult CT Examinations, Radiology 2017, ahead of print. (<http://pubs.rsna.org/doi/abs/10.1148/radiol.2017161911?journalCode=radiology>)

# CT and other modalities

- Current efforts to expand DIR to other modalities



## CLINICAL STUDY

### **The American College of Radiology Fluoroscopy Dose Index Registry Pilot: Technical Considerations and Dosimetric Performance of the Interventional Fluoroscopes**

Kevin A. Wunderle, PhD, A. Kyle Jones, PhD,  
Shalmali Dharmadhikari, PhD, Xinhui Duan, PhD, Don-Soo Kim, PhD,  
Usman Mahmood, MS, Steve D. Mann, PhD, Jeffery M. Moirano, MS,  
Rebecca A. Neill, MS, and Alan H. Schoenfeld, MS

# Research capabilities of DIR

## CT Volumes from 2,398 Radiology Practices in the United States: A Real-Time Indicator of the Effect of COVID-19 on Routine Care, January to September 2020


Matthew S. Davenport, MD<sup>a</sup>, Tom Fruscello, MBA<sup>b</sup>, Mythreyi Chatfield, PhD<sup>b</sup>, Stefanie Weinstein, MD<sup>c</sup>, William F. Sensakovic, PhD<sup>d</sup>, David B. Larson, MD, MBA<sup>e</sup>

### Abstract

**Purpose:** To determine the effect of coronavirus disease 2019 (COVID-19) on CT volumes in the United States during and after the first wave of the pandemic.

**Methods:** CT volumes from 2,398 US radiology practices participating in the ACR Dose Index Registry from January 1, 2020, to September 30, 2020, were analyzed. Data were compared to projected CT volumes using 2019 normative data and analyzed with respect to time since government orders, population-normalized positive COVID-19 tests, and attributed deaths. Data were stratified by state population density, unemployment status, and race.


**Results:** There were 16,198,830 CT examinations (2,398 practices). Volume nadir occurred an average of 32 days after each state-of-emergency declaration and 12 days after each stay-at-home order. At nadir, the projected volume loss was 38,043 CTs per day (of 71,626 CTs per day; 53% reduction). Over the entire study period, there were 3,689,874 fewer CT examinations performed than



**Fig 1.** Nationwide relationship between CT imaging volume (blue lines; 7-day moving averages) and coronavirus disease 2019 (COVID-19)-attributed deaths (top, gray bars, normalized per 100,000 people), COVID-19 positive tests (middle, orange bars, normalized per 100,000 people), and state government executive orders (bottom). Data are from January 1, 2020, to September 30, 2020. The “notches” in CT volume during the months of May, July, and September correspond to 3-day Memorial Day, July 4th, and Labor Day weekends, and also occurred in 2019 (Fig. 2).

# The Swiss National CT Dose Registry – proof of concept project


- Proof-of-concept project run by Kantonsspital Aarau (Prof Schindera, Dr. Lima and Mr Sigrist) and funded by Federal Office of Public Health




- Aimed to evaluate the viability of such project including implementation and running cost, technical aspects and development of a prototype.

# Proof of Concept Stage - Platform

- Implementation of core functionalities
  - Data submission (Interface for connecting institutes) and DMS
  - Providing tailored reports
  - Data cleansing process
  - Basis for data protection (Discussions, procedure mechanisms)
- Focus on view
  - Data analysis for participating institutes:  
Radiologists/Technologist/Med. Phy.
- Incorporating state-of-the-art web technologies
  - Goal: Using the PoC as foundation for the realisation




## Current Statistics



### Diagnostic Reference Levels (75p)

Protocol	National DLP 2018 [mGy-cm]	Current Registry DLP [mGy-cm]	National CTDI 2018 [mGy]	Current Registry CTDI [mGy]
Head	890.0	887.3	51.0	51.3
Sinuses	420.0	423.8	25.0	23.6
Low Dose Sinuses	90.0	86.3	6.0	6.5
Neck	410.0	407.6	16.0	15.4
Angio (Carotid)	360.0	355.7	11.0	10.3
Chest	250.0	247.4	7.0	7.0
Angio (Pulmonary Embolus)	300.0	297.3	8.0	7.8
Abdomen-Pelvis	540.0	529.6	11.0	11.1

# Data analysis



15 Protocols  
(as per  
DRLs)


DLP and  
CTDi<sub>vol</sub>

### About

The diagram on the left shows the median Dose Length Product (DLP) and Computed Tomography Dose Index (CTDI) of each device for the selected protocol. You can change the protocol and metric by using the menus at the bottom. The blue lines indicate the national Diagnostic Reference Levels (DRL) (25. and 75. percentiles). The device names are anonymised for data protection reasons, nevertheless you can highlight the devices of your own institute or enable a colour coding of the manufacturers by using the corresponding switches. Moreover, the devices can be filtered to show only the models owned by your institute.


# Highlight own data

Own data



# By Vendors

Different Vendors




### About

The diagram on the left shows the median Dose Length Product (DLP) and Computed Tomography Dose Index (CTDI) of each device for the selected protocol. You can change the protocol and metric by using the menus at the bottom. The blue lines indicate the national Diagnostic Reference Levels (DRL) (25. and 75. percentiles). The device names are anonymised for data protection reasons, nevertheless you can highlight the devices of your own institute or enable a colour coding of the manufacturers by using the corresponding switches. Moreover, the devices can be filtered to show only the models owned by your institute.

# Filtering just same vendors

Same Vendors as local devices



# Conclusion

- There is a legal requirement to optimise patient exposures.
- Dose Management Systems, or Software Registries, are one of the possible tools to help teams to implement optimisation strategies in their workflow.
- Most solutions present similar features (connectivity, analysis, reporting) but other features are specific to some vendors.
- Hospitals are not required to have a commercial software but normally in-house or opensource option require extra (human) resources.
- **Lastly, but not least, the analysis of these dose registry software will be/is as good as the data it has inside – having such solutions is only the begging of the work**

Thank you for listening.