



PAPER • OPEN ACCESS

Proposed DRLs for mammography in Switzerland

To cite this article: Laura Dupont *et al* 2024 *J. Radiol. Prot.* **44** 021512

View the [article online](#) for updates and enhancements.

You may also like

- [Experimental examination of radiation doses from cardiac and liver CT perfusion in a phantom study as a function of organ, age and sex](#)
Denise Bos, Britta König, Sebastian Blex *et al.*
- [IPEM topical report: the first UK survey of dose indices from radiotherapy treatment planning computed tomography scans for adult patients](#)
Tim J Wood, Anne T Davis, James Earley *et al.*
- [Evaluation of dose radiation on x-ray radiography](#)
Z Arifin, E Hidayanto, B Rahayuningsih *et al.*



PAPER

Proposed DRLs for mammography in Switzerland

OPEN ACCESS

RECEIVED
4 September 2023REVISED
17 January 2024ACCEPTED FOR PUBLICATION
26 March 2024PUBLISHED
16 May 2024

Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](https://creativecommons.org/licenses/by/4.0/).

Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.



Laura Dupont^{1,*} , Christoph Aberle² , Diomidis Botsikas¹, Michael Ith³, Thiago Viana Miranda Lima⁴ , Roman Menz², Pascal Monnin⁵, Pierre-Alexandre Poletti¹, Stefano Presilla⁶, Alexander Schegerer⁷, Liana-Cristina Stoica¹, Philipp Trueb⁸ and Marta Sans Merce¹ 

¹ Geneva University Hospital, Geneva, Switzerland

² Basel University Hospital, Basel, Switzerland

³ Institute of Diagnostic, Interventional and Pediatric Radiology, Inselspital, Bern University Hospital, University of Bern, Bern, Switzerland

⁴ Lucerne Cantonal Hospital, Lucerne, Switzerland

⁵ Vaud University Hospital Centre, Lausanne, Switzerland

⁶ Lugano Regional Hospital, Lugano, Switzerland

⁷ Hirslanden Clinic, Zurich, Switzerland

⁸ Federal Office of Public Health, Bern, Switzerland

* Author to whom any correspondence should be addressed.

E-mail: laura.dupont@hcuge.ch

Keywords: diagnostic reference levels, mammography, radiation protection, Switzerland

Abstract

The aim of this study is to propose diagnostic reference levels (DRLs) values for mammography in Switzerland. For the data collection, a survey was conducted among a sufficient number of centres, including five University hospitals, several cantonal hospitals, and large private clinics, covering all linguistic regions of Switzerland to be representative of the clinical practice. The data gathered contained the mean glandular dose (MGD), the compressed breast thickness (CBT), the mammography model and the examination parameters for each acquisition. The data collected was sorted into the following categories: 2D or digital breast tomosynthesis (DBT) examination, craniocaudal (CC) or mediolateral oblique (MLO) projection, and eight categories of CBT ranging from 20 mm to 100 mm in 10 mm intervals. A total of 24 762 acquisitions were gathered in 31 centres on 36 mammography units from six manufacturers. The analysis showed that the data reflects the practice in Switzerland. The results revealed that the MGD is larger for DBT than for 2D acquisitions for the same CBT. From 20–30 mm to 90–100 mm of CBT, the 75th percentile of the MGD values obtained increased from 0.81 mGy to 2.55 mGy for 2D CC acquisitions, from 0.83 mGy to 2.96 mGy for 2D MLO acquisitions, from 1.22 mGy to 3.66 mGy for DBT CC acquisitions and from 1.33 mGy to 4.04 mGy for DBT MLO acquisitions. The results of the survey allow us to propose Swiss DRLs for mammography according to the examination type (2D/DBT), projection (CC/MLO) and CBT. The proposed values are very satisfactory in comparison with other studies.

List of abbreviations

CBT	compressed breast thickness
CC	craniocaudal
CR	computed radiography
DBT	digital breast tomosynthesis
DMS	dose management system
DR	digital radiography
DRL	diagnostic reference level
FFDM	full field digital mammography
MGD	mean glandular dose
MLO	mediolateral oblique

1. Introduction

Breast cancer is a highly prevalent cancer among women. In 2020, it was the most common cancer worldwide but was only the fifth on the list of the most common causes of cancer deaths; however, early diagnostics and screening via mammography examinations help to reduce cancer mortality [1].

In Switzerland, breast cancer affects around 6300 women each year, and is still the leading cause of death from cancer with approximately 1400 women dying from it annually [2].

Mammography is the basic examination of the breasts that can detect changes in breast tissue, diagnose and manage patients having breast disorders. On top of the diagnostic mammography performed for a suspicious breast tissue modification or appearance, a screening programme in Switzerland provides women over 50 the opportunity to have a mammography every two years.

In 2018, screening mammography represented 0.97% of the total x-ray examinations in Switzerland corresponding to 0.3% of the dose contribution of the x-ray modalities. For what concerns diagnostic mammography examinations, these represented 1.74% of the x-ray examinations, corresponding to 0.6% of the total dose of the x-ray modalities [3].

Breast tissue is known to be a highly radiosensitive organ [4–6]. Thus, it is crucial to optimize its exposure during diagnostic and screening mammograms. Unlike other diagnostic examinations, for mammography, the European Commission has established acceptable and achievable values of MGD per CBT ranging from 21 to 90 mm for screening examinations [7]. The concept of DRLs is recognized internationally as an important tool to optimize patient's exposure [8]. DRLs are used as benchmark values that help ensure the optimization of radiation dose in medical imaging procedures. By setting these reference levels, healthcare providers can consistently monitor and manage the radiation exposure delivered to patients during mammography. Comparing MGD values with national or international DRLs is a good way to get an overview of the practice. DRLs should be viewed not as limits, but as dose indicators, establishing a DRL will lead to optimisation when necessary. According to ICRP 135 [8] a DRL value is defined as the 75th percentile of the distribution of the median values of the participating institutes obtained by the survey.

In Switzerland, even though DRLs exist for almost all modalities [9], no DRL has yet been established for mammography.

The aim of the present study is to establish national DRLs for mammography examinations in Switzerland, based on a survey performed in 31 Swiss centres, including university hospitals, canton hospitals and private clinics, to obtain a representative overview of the practices. Data was collected for 2D full field digital mammography (FFDM which will be referred to as 2D in this article for simplicity) and DBT mammography examinations with a wide variety of parameters, such as different CBT, projections, and digital mammography devices.

2. Materials and methods

2.1. Data collection

The data collection was organized by contacting the medical physicists from the 5 University hospitals in Switzerland and those working in the different canton hospitals and large private clinics to cover all regions of Switzerland. In total, we collected data from 31 of the 206 (15%) institutes in Switzerland, and from 36 of the 252 (14%) digital mammography units in the country. The data was collected for 2D and DBT mammography examinations including both diagnostic and screening mammograms.

Several parameters influence the MGD of a mammogram. Different MGD are delivered for large CBTs [10, 11], for DBT examinations compared to 2D exams, and for MLO (mediolateral-oblique) compared to CC projection. Therefore, the questionnaire specifically requested 20 acquisitions per each 10 mm thickness interval and per projection, for 2D and DBT examinations.

The breast is composed of glandular and adipose tissues, each in different proportions specific to each individual which change with the breast thickness and age of the patient. The proportion of glandular/adipose tissue determines the breast density, which influences the MGD [12]. Glandularity is not easily accessible or exportable via the DMS and was not evaluated in this study. Therefore, some studies conducted on DRLs in mammography present results according to age categories in addition to other categories (2D/DBT), projections, CBT, etc [10].

2.2. Questionnaire

A questionnaire for data collection was sent to the medical physicists in early October 2021. Data was collected until March 2022.

The first part of the questionnaire consisted of general questions about the mammography unit (manufacturer and model of the system) and the centre (name of the institute, contact details of the

correspondent). One questionnaire per unit had to be provided. Instructions for completing the questionnaire accurately were also provided.

Requirements for the data were the following:

- Only data from female patients were gathered.
- It was aimed at a minimum of 20 acquisitions per each category of CBT interval (eight categories, from 20 to 100 mm thickness in 10 mm intervals), projection (CC and MLO) and for 2D and DBT separately.
- If 20 acquisitions were not feasible, respondents were requested to provide as much data as possible for each category, and the units were still included in the analysis.
- The provided data had to be the most recent consecutive data.

The second part of the questionnaire was the one that had to be filled in with the following data for each acquisition provided:

- acquisition type (2D/DBT)
- projection (CC/MLO)
- CBT
- MGD displayed by the mammography unit.

3. Data analysis

The analysis was performed with data arranged according to the following categories:

- 2D/DBT
- projections CC/MLO
- CBT.

Regarding the dosimetric quantity to use for establishing DRLs, the ICRP 135 [8] provides three different possibilities: Entrance-surface air kerma ($K_{a,e}$), Incident air kerma ($K_{a,i}$), and MGD, but suggests to use MGD, as also proposed in other publications [12, 13]. For this reason, and because MGD represents the mean absorbed dose in the breast glandular tissue and is related to the radiological risk of induced cancer, we decided to use MGD as dosimetric quantity in this study. This value, determined by mammography system, can be subject to a range of uncertainty, dependent on the system and the dose model used for calculation [14–17].

The proposed DRLs were defined for single acquisitions. DRLs for single acquisitions directly allow evaluating how well the acquisition parameters are optimized. DRLs for whole examinations would depend also on the number and type of acquisitions which is rather a question of justification than optimization.

Data was received from 31 centres, for a total of 36 devices. The total number of acquisitions collected was 24 762, 14 925 for 2D mammography (table 1) and 9837 for DBT exams (table 2), for both CC and MLO projections and for each 10 mm CBT interval. The minimum of 20 requested acquisitions could not be reached for the extreme categories of CBT for each mammography unit.

4. Results

Our first step was to evaluate whether the mammography units for which we collected the data were representative of the models used in Switzerland. In total, we analysed data from 14% of the mammography units installed in Switzerland. These mammography units represent more than 87% of the most commonly used models, which are the models for which 10 or more units are installed in Switzerland. For DBT examinations, we collected data from 4 out of the 7 models that can perform these examinations.

According to ICRP 135 [8], a DRL value is defined as the 75th percentile of the distribution of the medians of the datasets obtained through a survey. For each unit, the median MGD was therefore calculated for 2D/DBT exams, for CC/MLO projections and for each CBT interval. The distributions of the median MGD are depicted as boxplots in figure 1 for 2D mammograms and figure 2 for DBT mammograms.

The 75th percentile of the MGD values obtained for 2D and CBT between 20 mm and 100 mm are comprised between 0.81 mGy–2.55 mGy for CC and between 0.83 mGy–2.96 mGy for MLO. For DBT, the values are comprised between 1.22 mGy–3.66 mGy for CC and 1.33 mGy–4.04 mGy for MLO. MGD are higher for DBT than for 2D, and for MLO than for CC.

The substantial difference in MGD values between the lowest and highest CBT intervals shows the importance of analysing the data for different CBT values. The steady increase of the 75th percentile MGD

Table 1. Number of acquisitions provided for each mammography device model, for 2D, for both CC/MLO projections, and 10 mm CBT intervals.

Models	N units	CC										MLO										Total
		20–30 mm	30–40 mm	40–50 mm	50–60 mm	60–70 mm	70–80 mm	80–90 mm	90–100 mm	100 mm	30–40 mm	40–50 mm	50–60 mm	60–70 mm	70–80 mm	80–90 mm	90–100 mm					
Hologic—Selenia dimensions	6	69	188	335	450	446	198	54	24	117	241	357	459	542	315	114	40	3949				
Hologic—3Dimensions	2	25	74	82	143	127	60	39	6	29	69	80	127	131	81	52	18	1143				
Siemens—Mammomat inspiration	11	204	203	203	203	205	184	187	106	136	136	137	137	124	117	111	101	2494				
Siemens—Mammomat revelation	3	121	296	582	841	753	386	102	26	11	30	108	252	253	158	41	20	3980				
IMS	5	100	100	100	100	100	100	97	31	100	100	100	100	100	100	100	77	1505				
Giotto—Giotto class	2	18	26	39	64	58	33	15	6	21	23	38	55	60	41	25	4	526				
Electric—senographe Essential	1	11	8	12	13	8	9	9	7	9	12	8	7	12	11	11	13	160				
Philips—Microdose mammo-graphy L30	1	8	8	12	9	10	11	8	7	12	12	8	11	10	9	12	13	160				
Philips—Microdose L50	1	11	8	13	8	10	10	8	1	9	12	7	12	10	10	12	0	141				
Philips—Mammo diagnost DR	3	40	41	57	56	65	55	54	15	44	71	70	87	85	63	43	21	867				
Fujifilm—Amulet FDR	35	607	952	1435	1887	1782	1046	573	229	488	706	913	1247	1327	905	521	307	14925				

Table 2. Number of acquisitions provided for each mammography device model, for DBT, for both CC/MLO projections, and 10 mm CBT intervals.

Models	N units	CC										MLO										Total
		20–30 mm		30–40 mm		40–50 mm		50–60 mm		60–70 mm		70–80 mm		80–90 mm		90–100 mm		100 mm				
		30 mm	20 mm	40 mm	30 mm	50 mm	40 mm	60 mm	50 mm	70 mm	60 mm	80 mm	70 mm	90 mm	80 mm	100 mm	90 mm	100 mm				
Hologic—Selenia dimensions	4	44	72	103	72	109	62	7	1	62	89	111	146	136	100	64	22	1245				
Hologic—3Dimensions	2	104	378	699	1030	648	251	64	13	116	377	610	907	741	374	123	51	6486				
Siemens—Mammomat inspiration	5	21	25	39	40	40	30	19	1	61	76	81	96	100	82	66	34	811				
Siemens—Mammomat revelation	1	—	—	—	—	—	—	—	—	20	20	20	20	20	20	20	20	160				
IMS	4	77	80	80	80	80	80	67	37	64	77	80	80	80	80	59	34	1135				
Giotto—Giotto class	16	246	555	921	1267	877	423	157	52	323	639	902	1249	1077	656	332	161	9837				




Figure 1. Boxplots of median values of MGD of all mammography units, for 2D acquisitions, CC and MLO projections, for all CBT intervals.




Figure 2. Boxplots of median values of MGD of all mammography units, for DBT acquisitions, CC and MLO projections, for all CBT intervals.

Table 3. Proposed DRLs values for mammography examinations in Switzerland (MGD per acquisition).

CBT (mm)		CC								MLO							
		20–30	30–40	40–50	50–60	60–70	70–80	80–90	90–100	20–30	30–40	40–50	50–60	60–70	70–80	80–90	90–100
2D	MGD	0.81	0.90	1.03	1.31	1.54	1.86	2.21	2.55	0.83	0.90	1.03	1.28	1.64	2.12	2.23	2.96
DBT	(mGy)	1.22	1.22	1.47	1.85	2.35	3.03	3.51	3.66	1.33	1.42	1.52	1.89	2.34	3.04	3.57	4.04

values from CBT interval to CBT interval indicates that the collected dataset was sufficiently large and that the 10 mm CBT intervals were large enough to calculate robust CBT-specific DRL values.

The advantage of considering the medians of each mammography unit is that they all have the same statistical weight in the final results. It is important that each unit has an equivalent weight compared to the others, otherwise the units that have provided more data will have a greater impact on the final results, which will no longer be representative of all the units. We therefore consider this method as the appropriate one to propose DRLs.

4.1. DRLs proposal

The 75th percentile of the distribution of the median MGD is presented in table 3 for 2D/DBT, for each projection (CC/MLO) and for each CBT interval. These results represent the proposed DRLs for Switzerland.

5. Discussion

Thanks to this survey, we have been able to propose a DRL value for 2D mammography acquisitions and DBT, for each 10 mm interval of CBT ranging from 20 to 100 mm, and for both CC and MLO projections. These results were computed by calculating the 75th percentile of the medians from all the mammography units that provided data. These results are proposed to become the first DRLs for mammography in Switzerland.

The comparison between our results with those from other countries should be made very carefully since different methodologies have been used, some made their DRLs using patient survey, and some using PMMA phantoms [10, 13, 18–22]. Therefore, we have chosen to compare our results with those obtained by countries that applied the same methodology (i.e. patient survey). When considering our study alongside others, it is crucial to acknowledge divergent statistical approaches in establishing DRLs. Unlike our use of the 75th percentile and median, some studies opt for the 95th percentile and mean values. These methodological distinctions may contribute to observed differences in reported dose levels, warranting careful interpretation when comparing findings across studies.

Some countries have, like us, chosen to separate their results by CBT intervals as shown in table 4. The New South Wales study shows results for separated CBT intervals of 10 mm likewise our results. Their range is larger, from 20 mm to 110 mm, while we have considered a maximum CBT of 100 mm [21]. The New South Wales study considers different detector technologies, but, since we only have DR systems, we have compared our results only with the DR ones. The values they obtained for DR vary from 0.97 to 2.63 mGy generally higher than ours for all CBT intervals and projections, except for 70–80 mm CBT interval for MLO projection, and for 90–100 mm interval for MLO projection, where our values are slightly higher. For what concerns the results obtained for the CR technology, they are much higher than our results.

The Scotland study also presents results for DR and CR, for several larger CBT intervals of 20 mm and include values of CBT below 30 mm and above 80 mm (for DR only) [19]. Their DR results vary from 0.73 to 2.81 mGy, generally higher than ours, except for the CBTs below 30 for which their value is lower than our 20 mm to 30 mm result. For CR, their values vary from 1.59 to 3.22 and are higher than ours.

In the study from Turkey, the results presented are separated in the same CBT and projection categories than our study, adding two age categories for patients aged 40–49 and 50–64 [10]. The calculated values vary from 2.00 to 2.60 mGy for patients aged 40–49 CC projection, and from 2.40 to 3.5 mGy for MLO projection. Results vary from 2.00 to 3.00 mGy for CC projection and from 2.30 to 4.00 for MLO projection for patients aged 50–64. The results from this study are generally higher than ours, except for the 90–100 mm CBTs for CC projection for both age categories, where our values are higher. An interesting point is the introduction of the age as a category for establishing DRLs. As mentioned earlier, breast density (glandularity) has an influence on the delivered dose and it is indeed changing with age. Nevertheless, the key parameter remains glandularity rather than age. Unfortunately, glandularity is not an easily accessible parameter nor exportable via the DMS.

Table 4. Results for 2D/DBT and CC/MLO projections and for 10 mm CBT intervals (MGD per acquisition), in comparison with other studies.

			CBT (mm)								
			20–30	30–40	40–50	50–60	60–70	70–80	80–90	90–100	100–110
MGD (mGy)	Our study	2D—CC	0.81	0.9	1.03	1.31	1.54	1.86	2.21	2.55	—
		DBT—CC	1.22	1.22	1.47	1.85	2.35	3.03	3.51	3.66	—
		2D—MLO	0.83	0.9	1.03	1.28	1.64	2.12	2.23	2.96	—
		DBT—MLO	1.33	1.42	1.52	1.89	2.34	3.04	3.57	4.04	—
	New South Wales	2D	0.97	1.12	1.3	1.65	2.35	2.08	2.34	2.63	3.31
	Scotland	2D	0.73		1.08	1.51		2.16	2.81	—	—
	Turkey (Age 40–49)	2D—CC	2	2.2	2.2	2.3	2.6	2.6	2.3	2.5	—
	Turkey (Age 40–49)	2D—MLO	3	2.4	2.5	2.6	2.8	3.2	3.3	3.5	—
	Turkey (Age 50–64)	2D—CC	2.6	2.1	2.2	2	2.5	2.6	3	2.2	—
	Turkey (Age 50–64)	2D—MLO	3.1	2.4	2.5	2.3	2.5	2.9	3.3	4	—

Table 5. Results for all data pooled for 2D/DBT and CC/MLO projections and for a mean CBT value (MGD per acquisition), in comparison with other studies. ‘Std’ is the standard deviation of the set of data.

MGD (mGy)		2D		DBT	
		CC	MLO	CC	MLO
Our study	Median	1.10	1.18	1.61	1.75
	75th perc.	1.29	1.42	1.79	1.91
	95th perc.	1.55	1.84	1.94	2.10
	Std	0.26	0.34	0.22	0.31
	Mean CBT (mm)	56	58	54	56
France	Median		1.70		2.30
	Mean CBT (mm)		56		57
Ghana	75th perc.		2.00		—
	95th perc.		4.20		—
	Mean CBT (mm)		60		—

Other countries have set their DRLs values for an average CBT value, some separating for CC and MLO projections. To compare our work with their results, we use the values obtained by pooling the data for 2D and DBT separately, for CC and MLO projections, for a mean CBT value. The results for our data when proceeding this way are presented in table 5, in comparison with two other studies with similar methodologies.

DRLs proposed by IRSN in 2021 [18] were 1.7 mGy for 2D for a mean CBT of 56 mm, and 2.3 mGy for DBT for a mean CBT of 57 mm. In our study, DRLs for 2D mammograms are 1.29 mGy for an average CBT of 56 mm and 1.42 mGy for an average CBT of 58 mm for CC and MLO respectively. DRLs for DBT examinations are 1.79 mGy for an average CBT of 54 mm and 1.91 mGy for an average CBT of 56 mm for CC and MLO respectively. The values of the French report and ours, when obtained using the same method, are quite similar. Slightly higher values have been obtained in Ghana for the 2D, with a DRL set at 75th percentile with a MGD of 2 mGy for a CBT 60 ± 5 mm [22].

In conclusion, our results are in good agreement with those of other countries when using a similar method.

6. Limitations

The findings of this study have to be seen in light of some limitations. First, during data collection, some institutes, mainly small institutes, were unable to participate, as they did not have a DMS. Gathering data without DMS can be very time-consuming and tedious. This greatly limited the number of institutes willing to participate in our study, and therefore the amount of data.

We encountered a second limitation during data collection. We had asked the participating institutes to send us, if possible, 20 data per category of 10 mm intervals of CBT. However, cases for extreme CBTs (very small or very large) are less common than intermediate values. Many centres were therefore unable to collect the 20 data requested for these categories.

We also had to neglect the influence of age and/or breast glandularity, which can have a strong impact on the dose delivered. To consider the influence of these factors, we should have added different categories of glandularity, which would have enormously increased the amount of data expected for each centre. Moreover, glandularity is a not easily accessible data via the DMS, and collecting this information would therefore have been very time-consuming.

Furthermore, while our study relies on the MGD determined by the mammography system, it is important to acknowledge the inherent uncertainty associated with this quantity. The accuracy of the recorded MGD values is contingent upon factors such as the mammography system in use and the dose model employed for calculation. These inherent uncertainties highlight a limitation in the precise quantification of MGD levels and emphasize the need for cautious interpretation of the results in the context of these potential variations.

Finally, variations in dose levels across different manufacturers' equipment could contribute to observed differences in our study. Notably, the data, particularly in DBT, is predominantly based on one supplier. Caution should be exercised when extrapolating findings to other manufacturers. Recognizing this variability in equipment is essential for a nuanced interpretation of our results.

7. Conclusion

To propose DRL values for mammography examinations performed in Switzerland we followed the recommendations established by the International Commission on Radiological Protection [8]. To this end, we performed a survey allowing us to collect data from 36 mammography units installed all over the country. The data, categorized into different categories such as examination type (2D or DBT), projection (CC or MLO) and 8 different CBT categories of 10 mm width, ranging from 20 mm to 100 mm was analysed. The data is representative of the practice in Switzerland since the most frequently installed models are represented. The MGD is larger for a DBT acquisition than for 2D, it increases as the CBT increases and has higher values for MLO than CC. Finally, DRLs were proposed as a function of the examination type (2D/3D), projection (CC/MLO) and CBT. The proposed values compare well to those obtained in the literature.

Data availability statement

The data cannot be made publicly available upon publication because no suitable repository exists for hosting data in this field of study. The data that support the findings of this study are available upon reasonable request from the authors.

Acknowledgments

The research leading to these results received funding from Federal Office of Public Health (FOPH).

Critical relevance statement

The first establishment of DRLs for mammography in Switzerland is the starting point for an optimized practice.

ORCID iDs

Laura Dupont  <https://orcid.org/0009-0004-6268-7482>

Christoph Aberle  <https://orcid.org/0000-0002-8763-4801>

Thiago Viana Miranda Lima  <https://orcid.org/0000-0002-3105-9270>

Marta Sans Merce  <https://orcid.org/0000-0002-8201-2720>

References

- [1] World Health Organization 2022 *Cancer* (available at: www.who.int/mediacentre/factsheets/fs297/en/) (Accessed 5 April 2022)
- [2] Dépistage du cancer du sein Swiss Cancer Screening 2023 Swiss cancer screening (available at: www.swisscancerscreening.ch/fr/geneve/depistage-du-cancer-du-sein) (Accessed 24 June 2023)
- [3] Viry A, Bize J, Trueb P R, Ott B, Racine D, Verdun F R and LeCoultré R 2021 Annual exposure of the Swiss population from medical imaging in 2018 *Radiat. Prot. Dosim.* **195** 1–7
- [4] National Research Council 2006 *Health Risks from Exposure to Low Levels of Ionizing Radiation: BEIR VII—Phase 2* (The National Academies Press)
- [5] The International Commission on Radiological Protection 2007 *Annals of the ICRP. Publication 103. The 2007 Recommendations of the International Commission on Radiological Protection* (SAGE Publications Inc.)
- [6] United Nations Scientific Committee on the Effects of Atomic Radiation 2008 *Effects of Ionizing Radiation* (United Nations)
- [7] Karsa L, Holland R, Broeders M, Wolf C, Perry N and Törnberg S (eds) (European Commission, Directorate-General for Health and Consumers) 2013 *European Guidelines for Quality Assurance in Breast Cancer Screening and Diagnosis* 4th edn, supplements (Publications Office)
- [8] ICRP 2017 Diagnostic reference levels in medical imaging. ICRP publication 135 *Ann. ICRP* **46**
- [9] OFSP 2022 *DRLs in Switzerland* (available at: www.bag.admin.ch/bag/fr/home/gesetze-und-bewilligungen/gesuche-bewilligungen/bewilligungen-aufsicht-im-strahlenschutz/informationen-fuer-medizinische-betriebe/diagnostische-referenzwerte-im-strahlenschutz.html) (Accessed 5 April 2022)
- [10] Parmaksız A, Ataç G K, Bulur E, Inal T and Alhan A 2020 Average glandular doses and national diagnostic reference levels in mammography examinations in Turkey *Radiat. Prot. Dosim.* **190** 100–7
- [11] Meybluma E, Gardavauda F, Daoa T-H, Fourniera V, Beaussarta P, Pigneura F, Baranesa L, Rahmounia A and Luciani A 2015 Breast tomosynthesis: dosimetry and image quality assessment on phantom *Diagn. Intervent. Imaging* **96** 931–9
- [12] Samara E T, Tsapaki V and Sramek D 2019 Dose management software implementation in mammography *Phys. Med.* **68** 88–95
- [13] European Commission 2014 *Diagnostic Reference Levels in Thirty-Six European Countries. Radiation Protection N° 180 Part 2*
- [14] Suleiman M E, Brennan P C and McEntee M F 2017 Mean glandular dose in digital mammography: a dose calculation method comparison *J. Med. Imaging* **4** 013502
- [15] Dance D R, Skinner C L and Carlsson G A 1999 Breast dosimetry *Appl. Radiat. Isot.* **50** 185–203
- [16] Boone J M 2002 Normalized glandular dose (DgN) coefficients for arbitrary x-ray spectra in mammography: computer-fit values of Monte Carlo derived data *Med. Phys.* **29** 869–75
- [17] Dance D R 1990 Monte Carlo calculation of conversion factors for the estimation of mean glandular breast dose *Phys. Med. Biol.* **35** 1211–9
- [18] IRSN 2021 AVIS IRSN N° 2021-00193 (available at: www.irsn.fr/FR/expertise/avis/2021/Pages/Avis-IRSN-Decembre-2021.aspx#Ykv5BNNBxTY) (Accessed 5 April 2022)
- [19] Weir A, Schofield K A and McCurrach A 2021 Setting Scottish diagnostic reference levels for mammography incorporating both craniocaudal and oblique projections between 30 and 80 mm *J. Radiol. Prot.* **41** 97–117
- [20] Suleiman M E, Brennan P C and McEntee M F 2015 Diagnostic reference levels in digital mammography: a systematic review *Radiat. Prot. Dosim.* **167** 608–19
- [21] Suleiman M E, McEntee M F, Cartwright L, Diffey J and Brennan P C 2017 Diagnostic reference levels for digital mammography in New South Wales *J. Med. Imaging Radiat. Oncol.* **61** 48–57
- [22] Dzidzornu E, Angmorther S K, Ofori-Manteaw B B, Aboagye S, Dzefi-Tettey K and Ofori E K 2021 Mammography diagnostic reference levels (DRLs) in Ghana *Radiography* **27** 611–6